



AEROSOL
MAGEE SCIENTIFIC

MEASURE BLACK CARBON CONTENT OF FILTER SAMPLES



OT21

SOOTSCAN™
TRANSMISSOMETER

Measure Black
Carbon content
of filter samples

KEY FEATURES

- 2-Wavelength operation: UV (370 nm) & IR (880 nm)
- Non-contact, non-destructive, non-contaminating
- Accepts 25, 37 or 47 mm diameter filters
- Analyzes glass fiber, quartz fiber, Teflon filters

APPLICATIONS

- Air Quality monitoring
- Analysis of historical archives
- Identification of "biomass burning" aerosols
- Personal exposure monitoring
- Stack & emission testing
- Climate change monitoring

AEROSOL MAGEE SCIENTIFIC

PRODUCT SPECIFICATIONS

MEASUREMENT PRINCIPLE

Measurement of the attenuation of transmitted light due to an aerosol deposit previously collected on the filter. Simultaneous analysis at 370 nm and 880 nm.

INTERPRETATION

Light absorption measurement at 880 nm interpreted as Black Carbon ('BC' also called 'Elemental Carbon - EC'). Measurement at 370 nm designated as UVPM, interpreted as a 'Brown Carbon' ('BrC') indicator of aromatic organic compounds such as are found in smoke from biomass burning.

DATA FORMAT

Attenuation ($100 * \ln[I_0/I]$) in 'ATN' units at two wavelengths.

SPECIFICITY

No other aerosol species absorbs light even 0.001 times as much as Black Carbon in the visible range.

INTERCOMPARISON

Light absorption measurements can be converted to BC or EC equivalents.

SENSITIVITY

1 ATN unit, equivalent to $0.06 \mu\text{g}/\text{cm}^2$ BC on filter. If collected on a 47 mm diameter filter at a 16.7 LPM flow rate for 24 hours, this represents a Limit Of Detection for BC of $0.075 \mu\text{g}/\text{m}^3$.

MEASUREMENT RANGE

Maximum loading of 125 'ATN' units is optimal, equivalent to an average BC concentration of $3.1 \mu\text{g}/\text{m}^3$ collected on a 47 mm diameter filter at 16.7 LPM for 24 hours. At higher 'ATN' values, the linearity may be impaired. The Data Reduction Template suggests other choices of filter collection area, sampling flow rate and sample collection time to permit accurate data over a very wide range of BC concentrations.

SOOTSCAN TRANSMISSOMETER OT21

SAMPLE MEDIUM

Quartz fiber, T60 Teflon coated borosilicate glass fiber, or Teflon membrane filters may be analyzed. For the use of other substrates contact us.

SAMPLE ANALYSIS AREA

Accommodates 25 mm, 37 mm, and 47 mm diameter filter media.

SAMPLE RETENTION

Analysis is non-destructive, non-contact and non-contaminating. Filters are unaffected for subsequent laboratory analysis.

VALIDATION

NIST-traceable neutral density filter kit option available.

DATA OUTPUT

Digital data available via rear RS-232 (COM) and USB port.

DISPLAY AND INTERFACE

2-line display screen with keypad.

PHISYCAL SPECIFICATIONS

- Dimensions (HxWxD): 102 x 270 x 274 mm
- Weight: 5.5 kg
- Electrical Power supply: 100 - 230VAC 50/60Hz (auto-switching)
- Temperature: 0-40°C

ACCESSORIES

- Optical Diffuser filters required for the analysis of samples collected on Teflon membranes: 47mm. pack of 25 (PN M8013)
- ND Filter Validation Kit set of 4 (PN M7475)

AEROSOLMAGEESCI.COM



Scan for more
information



Benchmark Monitoring

Australian Sales and Support.

Contact Benchmark Monitoring
for more information:

Email: sales@benchmarkmonitoring.com.au
Website: <http://www.benchmarkmonitoring.com.au/>

Mobile: +(614) 01666077
Office: +(612) 65721028

Address: US / 17 Enterprise Crescent,
McDougalls Hill NSW 2330 Australia